

# TESTEC ACTIVE DIFFERENTIAL PROBES

## TT-SX 9001 / TT-SX 9002



	TT-SX 9001	TT-SX 9002
Order Number	15601	15602
Attenuation Ratio	x10 / x100	x20 / x200
DC-Accuracy	±2%	±2%
Bandwidth (-3dB)	30MHz	30MHz
Max. Differential Input Voltage (DC + peak AC)	±70V @ x10 ±700V @ x100	±140V @ x200 ±1400V @ x200
Max. Common Mode Input Voltage (DC + peak AC)	±700V @ x10 ±700V @ x100	±1400V @ x20 ±1400V @ x200
Absolute Maximum Input Voltage (each side to ground)	1000V CAT III	
Safety Standard	IEC/EN 61010-031:2015 cTUVus certified – Certificate No. 72228575	
Power requirements	4 x AA batteries or USB-C cable (supplied)	

Distributed by:

**testoon**.COM  
The measurement website

99 rue Beranger 92320 Chatillon - France  
Tel. : +33 (0) 1 71 16 17 00  
E-mail: [contact@testoon.com](mailto:contact@testoon.com)  
[www.testoon.com](http://www.testoon.com)

Technical data subject to change  
© Testec Elektronik GmbH 2022

[info@testec.de](mailto:info@testec.de)  
[www.testec.de](http://www.testec.de)

**TESTEC**